

CALIBRATION CERTIFICATE

for
Kustom LaserCam 4 Speed Meter

Customer: Derbyshire Police **Serial Number:** LE2520

Certificate Number: 4670

Date of Calibration: 28 April 2021 **Next Calibration:** 27 April 2022

Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CD748BAF
Oscillator Frequency - 19.999955 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming - Vertical - Pass
 - Horizontal - Pass
Eye Safety - Pass

Range \pm 0.2 m

Actual	47.64 m	86.72 m	148.86 m
Indicated	47.6 m	86.7 m	148.9 m

Speed \pm 1 MPH

Actual	26 MPH	38 MPH	-55 MPH	-62 MPH	73 MPH
Indicated	26 MPH	38 MPH	-55 MPH	-62 MPH	73 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK


L Korcz
Technical Signatory

TRUVELO (UK) LTD