

CALIBRATION CERTIFICATE

for

Kustom LaserCam 4 Speed Meter

Customer: Derbyshire Police **Serial Number:** LE2520

Certificate Number: 6426

Date of Calibration: 22 September 2022 **Next Calibration:** 21 September 2023

Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Oscillator Frequency - 19.999943 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming - Vertical - Pass
 - Horizontal - Pass
Eye Safety - Pass

Range ± 0.2 m

| | | | |
|-----------|---------|---------|----------|
| Actual | 47.64 m | 86.72 m | 148.86 m |
| Indicated | 47.6 m | 86.7 m | 148.9 m |

Speed ± 1 MPH

| | | | | | |
|-----------|--------|---------|---------|--------|--------|
| Actual | 58 MPH | -72 MPH | -66 MPH | 25 MPH | 38 MPH |
| Indicated | 58 MPH | -72 MPH | -66 MPH | 25 MPH | 38 MPH |

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK


L Korcz
Technical Signatory

TRUVELO (UK) LTD