CALIBRATION CERTIFICATE

Kustom LaserCam 4 Speed Meter

Customer:

Derbyshire Police

Serial Number:

LE2795

Certificate Number:

6427

Date of Calibration:

22 September 2022

Next Calibration:

21 September 2023

Place of Calibration:

14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Oscillator Frequency

- 19.999822 MHz

Self Test Sequence

- Pass

External Controls

- Pass

- Pass

Aiming Horizontal

Vertical

- Pass

Eye Safety

- Pass

Range ± 0.2 m

Actual

47.64 m

86.72 m

148.86 m

Indicated

47.6 m

86.7 m

148.9 m

Speed ± 1 MPH

Actual

22 MPH

-75 MPH

31 MPH

44 MPH

-60 MPH

Indicated

22 MPH

-75 MPH

31 MPH

44 MPH

-60 MPH

This is to certify that the Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that any parts replaced during the calibration process are identical to those used in the type approved device.

For Truvelo UK

L Korcz

Technical Signatory

TRUVELO (UK) LTD

Tel:020 8847 4400 TRU-CAL-CER-009

email:info@truvelouk.com