

CALIBRATION CERTIFICATE

for

Kustom LaserCam 4 Speed Meter

Customer: Derbyshire Police **Serial Number:** LE2520
Certificate Number: 7393
Date of Calibration: 03 October 2023 **Next Calibration:** 02 October 2024
Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CD748BAF
Oscillator Frequency - 19.999952 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming - Vertical - Pass
Horizontal - Pass
Eye Safety - Pass

Range ± 0.2 m
Actual 47.64 m 108.69 m 148.86 m
Indicated 47.6 m 108.7 m 148.9 m

Speed ± 1 MPH
Actual 74 MPH 61 MPH -26 MPH -35 MPH 48 MPH
Indicated 74 MPH 61 MPH -26 MPH -35 MPH 48 MPH

The sample measurements in the calibration certify this device for use to the full extent of the instrument specification Range 3 m - 750 m, Speed 0 mph - 200 mph

The Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truvelo UK


L Korcz
Technical Signatory

TRUVELO (UK) LTD