

CALIBRATION CERTIFICATE

for

Kustom LaserCam 4 Speed Meter

Customer: Derbyshire Police **Serial Number:** LE2520
Certificate Number: 9231
Date of Calibration: 20 August 2025 **Next Calibration:** 19 August 2026
Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CD748BAF
Oscillator Frequency - 19.999954 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming - Vertical - Pass
 - Horizontal - Pass
Eye Safety - Pass

Range ± 0.2 m

Actual	47.64 m	108.69 m	148.86 m
Indicated	47.6 m	108.7 m	148.9 m

Speed ± 1 MPH

Actual	35 MPH	73 MPH	-26 MPH	-48 MPH	64 MPH
Indicated	35 MPH	73 MPH	-26 MPH	-48 MPH	64 MPH

The sample measurements in the calibration certify this device for use to the full extent of the instrument specification
Range 3 m - 750 m, Speed 0 mph - 200 mph

The Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truvelo UK

L Korcz
Technical Signatory

TRUVELO (UK) LTD